

**EC DECLARATION OF CONFORMITY**

---

**Medical Device Manufacturer:**

**James Leckey Design Ltd**  
19C Ballinderry Road  
Lisburn  
BT28 2SA  
United Kingdom

Phone: + 44 (0) 28 9260 0750  
Fax: + 44 (0) 28 9260 0799  
E-mail: [info@leckey.com](mailto:info@leckey.com)  
Internet: [www.leckey.com](http://www.leckey.com)

**James Leckey Design Ltd declares that the Firefly Upsee Walking Harness, in all sizes, configurations and colours, along with its accessories, conforms to the Essential Requirements of the Medical Device Directive 93 / 42 / EEC (as amended by 2007 / 42 / EC)**

**Medical Device Name** Firefly Upsee Walking Harness & accessories

**MDD Classification** Class 1 (Classification Rule 1)

**Conformity Assessment Route**

In accordance with article 11 of Council Directive 93 / 42 / EEC, and for the purpose of CE marking the Firefly Upsee Walking Harness and its accessories, James Leckey Design Ltd followed the procedure relating to the EC Declaration of Conformity set out in Annex VII of the Directive.

Conformity was demonstrated by converting Annex I (Essential Requirements) of the Directive into a checklist, enabling James Leckey Design Ltd to ensure that all of the essential requirements have been met.

The technical documentation for this product and supporting information for this Declaration of Conformity is held at our offices at the above address.

**Approved by:**  
**Gary O'Hagan**  
Managing Director



**Date:**

**6<sup>th</sup> January 2016**